

**EMC TEST REPORT**

PAPIMI 600 SP

Testreport Nr.: 0346806E

Amendment: \_\_\_\_\_

PAGE : 3 of 12

INTRACOM S.A.  
19.5 MARKOPOULO AVE. - PEANIA, ATTIKA  
Quality Assurance Department  
Emc & Safety Laboratory

**CUSTOMER INFORMATION:**

NAME :	EMC ΕΛΛΑΣ Α.Ε	ADDRESS :	Μισογύλιου 2-4
DEPT :		CITY :	Αθήνα, Αττική, - Τ.Κ: 11527

DESCRIPTION OF PRODUCT:	Magnetic Pulse Generator
NAME OF PRODUCT:	PAPIMI 600 SP
PRODUCT NUMBER:	
SERIAL NUMBER:	
TRADE MARK:	"ΤΖΑΜΠΑΣ ΧΡ.-ΚΑΤΩΠΟΔΗΣ ΑΠ.Ο.Ε"

DATE OF RECEIPT OF PRODUCT : 07-11-2003

EMISSIONS ENVIR.	IEC 60901-1-2				
IMMUNITY ENVIR.	IEC 60901-1-2				

R	T	TEST DESCRIPTION	STANDARD	DATE	TESTED BY	PASS/FAIL
		Conducted emissions, PS, 230VAC	EN 55022			
X	X	Radiated Emissions	EN 55011	12/11/2003	K.M	PASS
X	X	Radiated E.M. field immunity	EN 61000-4-3	12/11/2003	K.M	PASS
		HF conducted immunity on PS	EN 61000-4-6			
		HF conducted immunity on RF port	EN 61000-4-6			
		Surges on PS lines	EN 61000-4-5			
		Surges on telecom lines	EN 61000-4-5			
		Fast transient burst (EFT) on PS	EN 61000-4-4			
		Fast transient burst (EFT) on RF	EN 61000-4-4			
		Electrostatic discharges (ESD)	EN 61000-4-2			
		Voltage dips, short interrupt. on PS	EN 61000-4-11			

R : Required, T : Tested

ISSUED BY		APPROVED BY	
K.MARKAKIS		J.CHARALAMPAKIS	
14/11/2003		14/11/2003	
DATE	SIGNATURE	DATE	SIGNATURE