



E53.1:C3/FTR601_1_1

Ref. No L30. 00.163.08.01

page 1.

TEST REPORT

Test item description : PULSE GENERATOR
Applicant : EMC HELLAS
Manufacturer : Z ELECTRODYNAMICS
Type / serial number : PAPIMI 600SP
Date of receipt of test item : 04 / 02 / 2004
Date of issue of this test report : 25 / 02 / 2004
Number of test samples : 1
Testing laboratory : **LABOR S.A.**
 Reg. No : 12937/134, Ministry of development
 O.J. of Greek Government : 1021/20.11.97

Purpose of the tests and inspections

To verify the conformity of the product to the requirements of the following Standard, where applicable:

IEC 60601-1-1:1988 + A2:1995

Medical electrical requirements

Part 1: General requirements for safety

IEC 60601-1-1:2001

Collateral standard : Safety requirements for medical electrical systems

RESULT OF THE TEST : **This product is in compliance with the applicable requirements of the mentioned standard**

This report consists of 25 pages

Client



Tested by

Date / signature

Date / signature

This test report refers only to the item tested. Compliance of the test sample does not ensure compliance of the whole production of the manufacturer.

Only a full reproduction of this Report is allowed without the written approval of the testing laboratory